

2012

## Characterisation of the Acoustic Output of Three Sonoporation Drug Delivery Ultrasound Systems Using an Acoustic Radiation Force Balance

Tapihwa Mabvaro  
*Technological University Dublin*

Jacinta Browne  
*Technological University Dublin, jacinta.browne@tudublin.ie*

Follow this and additional works at: <https://arrow.tudublin.ie/scschphyart>



Part of the [Physics Commons](#)

### Recommended Citation

Mabvaro, T., Browne, J. : Characterisation of the Acoustic Output of Three Sonoporation Drug Delivery Ultrasound Systems Using an Acoustic Radiation Force Balance. *Physica Medica*, Vol. 28, no.4, 2012, p. 343. doi:10.1016/j.ejmp.2012.06.041

This Article is brought to you for free and open access by the School of Physics & Clinical & Optometric Science at ARROW@TU Dublin. It has been accepted for inclusion in Articles by an authorized administrator of ARROW@TU Dublin. For more information, please contact [arrow.admin@tudublin.ie](mailto:arrow.admin@tudublin.ie), [aisling.coyne@tudublin.ie](mailto:aisling.coyne@tudublin.ie).



This work is licensed under a [Creative Commons Attribution-NonCommercial-Share Alike 4.0 License](#)

## Characterisation of the acoustic output of three sonoporation drug delivery ultrasound systems using an acoustic radiation force balance

T MABVARO and J BROWNE

School of Physics, Dublin Institute of Technology, Ireland

In this study, the acoustic output from three sonoporation drug delivery ultrasound systems, namely Sonidel SP100 were characterised using an acoustic radiation force balance. Delivery of nucleic acid to a target tissue as a form of gene therapy is an important means of treating a variety of disorders, for example cardiovascular disease. One method of this type of treatment utilises ultrasound to stimulate cell membrane permeabilisation (sonoporation) for the purpose of transferring nucleic acids into cells. This approach offers advantages over competing technologies such as viral vectors, primarily as a result of its relatively non-invasive nature. However, the ultrasound sonoporation systems used to produce this permeability of the cell need to be fully characterised in order to provide the most effective treatment. Therefore, in this study acoustic output measurements were made for three SP 100 ultrasound systems and the output obtained from each system were compared to determine whether each of the three systems had similar performance. An absorbing target acoustic radiation force balance was used to obtain these measurements, the resolution of this system was 1 mN. The results from each system operating in continuous mode were compared using a t-test and the results were found to be similar at a significance level of  $p < 0.05$  and the re-testing demonstrated that three systems had repeatable outputs. Both the continuous and pulse wave mode settings were tested at different intensities and a representative of these results are presented in the figure below. As well as testing the manual setup, automated analysis was also carried out using a dedicated LABVIEW program. The results obtained for both the manual and automated set-up will be presented as well as the results for the different pulsed wave mode settings.